


DAB Chromogen/Substrate Bulk Pack (High Contrast)

Description: This product has been developed for applications that require high contrast between the chromogen and Hematoxylin counterstain. The resulting stain is a darker brown than standard DAB and somewhat more sensitive. 3,3'-Diaminobenzidine (DAB) is a widely used chromogen for immunohistochemical staining and immunoblotting. When in the presence of peroxidase enzyme, DAB produces a brown precipitate that is insoluble in alcohol. This product is available in a two component form consisting of a liquid, refrigerator stable DAB Chromogen and DAB Substrate (High Contrast). The use of liquid components reduces some risks associated with handling powders (ie. dust inhalation), and eliminates waste which often results from using tablets that require a predetermined final volume.

Uses/Limitations: Not to be taken internally.
For In-Vitro Diagnostic use only.
Histological applications.
Do not use if reagents become cloudy.
Do not use past expiration date.
Use caution when handling reagents.
Non-Sterile.



Control Tissue: Any well fixed paraffin embedded or frozen tissue section.


Ordering information regarding individual components on back page!

Kit Options:


<u>Item #</u>	<u>Component</u>	<u>Volume</u>	<u>Storage</u>
ACV250	DAB Substrate (High Contrast)	250 ml	2-8°C
	DAB Chromogen Concentrate	15 ml	2-8°C
ACV500	DAB Substrate (High Contrast)	500 ml	2-8°C
	DAB Chromogen Concentrate	30 ml	2-8°C
ACV999	DAB Substrate (High Contrast)	1000 ml	2-8°C
	DAB Chromogen Concentrate	60 ml	2-8°C

Precautions: Avoid contact with skin and eyes.
Harmful if swallowed.
Follow all Federal, State, and local regulations regarding disposal.

Storage: 2° C  8° C

 ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.


Emergo Europe
Prinsessegracht 20
2514 AP The Hague, The Netherlands

Procedure:

1. Rinse slide in deionized water prior to application of DAB mixture.
2. Combine 50-80µl of DAB Chromogen with 1ml of DAB Substrate and apply to tissue for 5 minutes.

NOTE: Range = 50-80µl of DAB Chromogen / 1 ml of DAB Substrate. 50µl /ml produces optimal staining quality, 80µl/ml produces maximum sensitivity. Combined mixture may be used for up to six hours.


OPTIONAL: For additional sensitivity, add 20µl of DAB Enhancer (cat.# ACM030, not included) to each 1ml of mixed solution.

3. Rinse slide using deionized water.
4. Apply mixture a second time and incubate for another 5 minutes.
5. Rinse slide in two changes of deionized water.
6. Counterstain as desired.
7. Dehydrate through graded alcohol, clear and coverslip in synthetic medium.

References:

1. Kuller, L., Schmidt, A., Mack, H., Durning, M., Birkebak, T., Reiner, M.T., Anderson, D.M., Morton, W.R., Agy, M.B. Systemic and Intestinal Immune Responses to HIV-2₈₇ Infection in Macaca nemestrina. *AIDS Research and Human Retroviruses*. August 2001, Volume 17(12): pages 1191-1204. Doi:10.1089/088922201316912790.
2. Schaefer, K.L., Brachwitz, K., et al. Expression Profiling of t(12;22) Positive Clear Cell Sarcoma of Soft Tissue Cell Lines Reveals Characteristic Up-Regulation of Potential New Marker Genes Including ERBB3. *Cancer Research*, May 15, 2004, Volume 64; page 3395. Doi: 10.1158/0008-5472.CAN-03-0809.
3. Oznurulu, Y., Celik, I., Sur, E., Telatar, T., Ozparlak, H. Comparative Skin Histology of the White New Zealand and Angora Rabbits: Histometrical and Immunohistochemical Evaluations. *Journal of Animal and Veterinary Advances*. 2009, Volume 8(9): pages 1694-1701.
4. Rivas-Arancibia, S., Guevara-Guzman, R., Lopez-Vidal, Y., Rodriguez-Martinez, E., Zanardo-Gomes, M., Angoa-Perez, M., Raisman-Vozari, R. Oxidative Stress Caused by Ozone Exposure Induces Loss of Brain Repair in the Hippocampus of Adult Rats. *Toxicological Sciences*. 2010, Volume 113(1): pages 187-197, doi.1093/toxsci/kfp252.
5. Gupta, P., Dhingra, K.K., Singhal, S., Khurana, N., Mandal, S., Mohta, A. Primary primitive neuroectodermal tumour (PNET) of testis: an unsuspected diagnosis. *Pathology*, February 2010, Volume 42, Issue 2: pages 179-181. Doi: 10.3109/00313020903494029.
6. Lee, J.H., Pyon, J.K., Kim, D.W., Lee, S.H., Nam, H.S., Kim, C.H., Kang, S.G., Lee, Y.J., Park, Y.J., Jeong, D.J., Cho, M.K. Elevated c-Src and c-Yes expression in malignant skin cancers. *Journal of Experimental & Clinical Cancer Research*. 2010, Volume 29: page 116. Doi: 10.1186/1756-9966-29-116.
7. Ozaltin, F., Ibsirlioglu, T., et al. Disruption of PTPRO Causes Childhood-Onset Nephrotic Syndrome. *The American Journal of Human Genetics*. July 15, 2011, Volume 89(1): pages 139-147. Doi: 10.1016/j.ajhg.2011.05.026.

 Storage: 2° C  8° C



ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.

EC REP
Emergo Europe
Prinsessegracht 20
2514 AP The Hague, The Netherlands

Instructions For Use ACV-IFU

Rev. Date: Nov. 4, 2021


Revision: 4


Page 3 of 3

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 – Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

Bulk Reagent Ordering Information and Current Pricing at www.scytek.com

Description:	Catalog #	Volume
DAB Chromogen Concentrate	ACB015	15 ml
	ACB030	30 ml
	ACB060	60 ml
	ACB125	125 ml
	ACB500	500 ml
DAB Substrate (High Contrast)	ACU250	250ml
	ACU500	500 ml
	ACU999	1000 ml

Storage: 2° C  8° C



ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.



Emergo Europe
Prinsessegracht 20
2514 AP The Hague, The Netherlands