

Instructions For Use

PBT-IFU

Rev. Date: Dec. 6, 2017

Revision: 4

Page 1 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

Phosphate Buffered Saline plus Tween 20 (20x) pH 7.6

Description: ScyTek Wash buffer formulations are an optimal formulation of pH stabilizers, salts and detergents

designed to effectively remove excess material from the microtiter plate wells without disrupting the ELISA binding reaction. By maintaining the proper buffering environment, unbound components can be washed away without suppressing antigen-antibody binding interactions, thereby reducing nonspecific

background and increasing the specific signal. Our Wash buffers do not contain hazardous

preservatives such as Azide or Mercury that may interfere with antibody-antigen binding interactions. For your convenience wash buffer is offered in a wide variety of formulations to meet the needs of your specific ELISA application. This product can be customized to meet the specific needs of your assay.

Inquire about custom vialing, labeling, kit assembly and drop shipping.

Contents: Phosphate buffered saline in reagent grade water. Tween 20 is added at a diluted concentration of 0.05%,

NaCl: 0.133M, K₂HPO₄: 0.0086M, KH₂PO₄: 0.0015M. Final pH of diluted buffer is 7.6±0.05. This product is

filtered to 0.2 microns.

Availability: REF # Volume Diluted Volume

125 ml	2.5 Liters
500 ml	10 Liters
1000 ml	20 Liters
10 Liters	200 Liters
20 Liters	400 Liters
	500 ml 1000 ml 10 Liters

Uses/Limitations: Not to be taken internally.

For In-Vitro Diagnostic use. For professional use. Histological applications.

Do not use if reagents become cloudy. Do not use past expiration date. Use caution when handling reagents.

Non-Sterile.

Ordering Information and Current Pricing at www.scytek.com



Precautions: Avoid contact with skin and eyes.

Harmful if swallowed.

Follow all Federal, State, and local regulations regarding disposal.

Storage: 18° C 25° C

ScyTek Laboratories, Inc. 205 South 600 West Logan, UT 84321 U.S.A. ϵ

Emergo Europe Molenstraat 15 2513 BH Hague, The Netherlands



Instructions For Use PRT_IFI I

Rev. Date: Dec. 6, 2017

Revision: 4

Page 2 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

Procedure:

- 1. Pour 50ml of Phosphate Buffered Saline plus Tween 20 (20x) pH7.6 in mixing flask and add water to final volume of 1000ml.
- 2. Stir briefly.

References:

- Adali G, Yorulmaz E, Ozkanli S, Ulasoglu C, Bayraktar B, Orhun A, Colak Y, Tuncer I. Serum concentrations of insulin-like growth factorbinding protein 5 in Crohn's disease. World Journal of Gastroenterology: WJG. 2013 Dec 21;19(47):9049.
- 2. Yun HY, Sung R, Kim YC, Choi W, Kim HS, Kim H, Lee GJ, You RY, Park SM, Yun SJ, Kim MJ. Regional distribution of interstitial cells of Cajal (ICC) in human stomach. The Korean Journal of Physiology & Pharmacology. 2010 Oct 1;14(5):317-24.
- 3. Cui Z, Mumper RJ. Chitosan-based nanoparticles for topical genetic immunization. Journal of Controlled Release. 2001 Aug 10;75(3):409-19.

Warranty:

No products or "Instructions For Use (IFU)" are to be construed as a recommendation for use in violation of any patents. We make no representations, warranties or assurances as to the accuracy or completeness of information provided on our IFU or website. Our warranty is limited to the actual price paid for the product. ScyTek Laboratories, Inc. is not liable for any property damage, personal injury, time or effort or economic loss caused by our products. Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining can cause inconsistent results. Variations in fixation and embedding or the inherent nature of the tissue specimen may cause variations in results. Endogenous peroxidase activity or pseudoperoxidase activity in erythrocytes and endogenous biotin may cause non-specific staining depending on detection system used.

Storage: 18° C



