

Instructions For Use AFH600-IFU

Rev. Date: 10/24/03

Revision: 2

Page 1 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Fax (435) 755-0015 - www.scytek.com

SensiTek Alk-Phos (Anti-Polyvalent) Ready-To-Use (500 slide)

Species of Origin: Goat

Antigen Specificity: Anti-Mouse, Rat, Rabbit, Guinea Pig IgG (H+L)

Preabsorbed Against: Human

Enzyme Conjugate: Alkaline Phosphatase

Chromogen Substrate: None provided

Procedure:

- 1. Deparaffinize and rehydrate tissue section.
- 2. Wash 2 times in buffer.
- 3. If required, incubate tissue in digestive enzyme.
- 4. Wash 4 times in buffer.
- 5. Apply Super Block (blue cap), and incubate for 5-10 minutes at room temperature to block nonspecific background staining. Note: Do not exceed 10 minutes or there may be a reduction in desired stain.
- 6. Wash 1 time in buffer.
- 7. Apply primary antibody and incubate according to manufacturer's protocol.
- 8. Wash 4 times in buffer.
- 9. Apply Biotinylated Link Antibody (yellow cap), and incubate for 15-20 minutes at room temperature.
- 10. Wash 4 times in buffer.
- 11. Apply Streptavidin/Alk-Phos Label (red cap), and incubate for 20 minutes at room temperature.
- 12. Rinse 4 times in buffer.
- 13. Apply chromogen intended for use with Alkaline Phosphatase.
- 14. Counterstain and coverslip.

- Troubleshooting Guide -

Overstaining:

- Concentration of the primary antibody was too high or the incubation time was too long.
- 2. Temperature during incubation was too high.
- 3. Incubation time with link antibody or streptavidin/enzyme label was too long.

The material contained herein is the property of ScyTek Laboratories, Inc. and is confidential. Unauthorized use or copying is prohibited.



Instructions For Use AFH600-IFU

Rev. Date: 10/24/03

Revision: 2

Page 2 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Fax (435) 755-0015 - www.scytek.com

Nonspecific Background Staining:

- 1. Rinsing between steps was inadequate.
- 2. Tissue was allowed to dry with reagents on.
- 3. Folds in tissue trapped reagents.
- 4. Tissue contains endogenous alkaline phosphatase.
- 5. Tissue contains endogenous biotin.
- 6. Antigen migrated in tissue.
- 7. Excessive tissue adhesive on slides.
- 8. Inadequate blocking with protein block.

Weak Staining:

- 1. Primary antibody concentration was too low or incubation time was too short.
- 2. Reagents are past their expiration date.
- Inadequate removal of wash water between steps, resulting in dilution of reagents.
- 4. Counterstain or mounting media were incompatible and dissolved the chromogen reaction product.
- 5. Room temperature was excessively cool.
- 6. The primary antibody does not recognize an antigen that survives fixation and embedding in high enough amounts.
- 7. Excessive incubation with protein block (Super Block).

No Staining:

- 1. Steps were inadvertently left out.
- 2. There is no antigen in the tissue.
- 3. The primary antibody is not of mouse, rat, rabbit or guinea pig origin.
- 4. Chromogenic substrate has been replaced with another that is not intended for use with alkaline phosphatase.