

Instructions For Use A00003-IFU-RUO

Rev. Date: Jan. 12, 2016

Revision: 4

Page 1 of 1

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Tel. (435) 755-9848 - Fax (435) 755-0015 - www.ScyTek.com

CD20, B-Cell Monoclonal Mouse Antibody (Ready-To-Use)

Catalog Number	Volume
A00003-0002	2 ml
A00003-0007	7 ml
A00003-0025	25 ml

Intended Use

For research use only.

Description

Immunogen: BALB/C mice were injected with human tonsil B cells.

Clone: L26

Isotype: IgG2a, Kappa.

Format: This antibody has been pretitered and quality controlled to work

on formalin-fixed paraffin-embedded as well as acetone fixed

cryostat tissue sections. No further titration is required.

Specificity: This antibody is specific to a 33 kD polypeptide present on the

majority of B cells in peripheral blood and lymphoid tissue and also with a minor component of 30 kD. No reactivity with other

hematopoietic cells has been observed.

Species Reactivity: Human, Others-not known

Positive Control: Tonsil
Cellular Localization: Cytoplasmic

Titer/Working Dilution: Ready-to-Use (no further dilution required)

Microbiological State: Nonsterile.

<u>Procedure</u>

- 1. Tissue Section Pretreatment (Required): Staining of formalin fixed, paraffin embedded tissue sections is enhanced by pretreatment with Citrate Plus (ScyTek catalog# CPI 500)
- 2. **Primary Antibody Incubation Time:** We suggest an incubation period of 30 minutes at room temperature. However, depending upon the fixation conditions and the staining system employed, optimal incubation should be determined by the user.
- 3. **Visualization:** For maximum staining intensity we recommend the "UltraTek HRP Anti-Polyvalent Lab Pack" (ScyTek catalog# UHP125, see IFU for instructions) combined with the "DAB Chromogen/Substrate Bulk Pack (High Contrast)" (ScyTek catalog# ACV500, see IFU for instructions


Materials and Reagents Required but not Provided

- 1. Control tissue and reagents
- 2. Xylene, graded alcohols, and deionized/distilled water
- 3. IHC detection system (Suggested: ScyTek Cat# UHP125 and ACV500)
- 4. Wash buffer for rinses (ScyTek Cat# TBT500)
- 5. Retrieval solution (ScyTek Cat# CPL500)
- 6. Hematoxylin counterstain and bluing reagent (ScyTek Cat# HMM500 and BRT500)
- 7. Mounting medium and coverslips

Note: ScyTek Laboratories has a wide range of IHC reagents and ancillaries that can be found at scytek.com.

Storage and Stability

Do not Freeze. Store at 2-8°C. Return to 2-8° immediately after use. Do not use after



CD20, staining on human tonsil tissue. Results were visualized using ScyTek's UHP125 detection system and DAB Chromogen/Substrate (High Contrast) ScyTek Cat# ACV500

expiration date printed on label. Verify visually that antibody has not been contaminated before use. Do not use if reagent becomes cloudy or precipitates.

Limitations

Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining can cause inconsistent results. Variations in fixation and embedding or the inherent nature of the tissue specimen may cause variations in results. Endogenous peroxidase activity or pseudoperoxidase activity in erythrocytes and endogenous biotin may cause non-specific staining depending on detection system used. This data sheet's recommendations and procedures were validated using ScyTek IHC reagents and may not be suitable for other detection systems.

Precautions

- 1. Contains Sodium Azide as a preservative (0.09% w/v), do not ingest. Sodium Azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. This product contains no hazardous material at a reportable concentration according to U.S. 29 CFR 1910.1200, OSHA Hazardous Communication Standard and EC Directive 91/155/EC.
- 2. Do not pipette by mouth.
- 3. Avoid contact of reagents and specimens with skin and mucous membranes.
- 4. Avoid microbial contamination of reagents or increased nonspecific staining may occur.
- 5. The user must validate any procedures and recommendations that differ from this data sheet
- 6. The SDS may be found at scytek.com

<u>References</u>

- 1. Ishii et al. Clin Exp Immunol 58: 183. 1984.
- 2. Mason et al. Am J Pathol 136: 1215, 1990.
- 3. Cartun et al. Am J Pathol 129: 415, 1987.

Warranty

No products or "Instructions For Use (IFU)" are to be construed as a recommendation for use in violation of any patents. We make no representations, warranties or assurances as to the accuracy or completeness of information provided on our IFU or website. Our warranty is limited to the actual price paid for the product. ScyTek Laboratories, Inc. is not liable for any property damage, personal injury, time or effort or economic loss caused by our products.



ScyTek Laboratories, Inc. 205 South 600 West Logan, UT 84321 U.S.A.

 ϵ

EC REP EmergoEurope (31)(0) 70 345-8570 Molenstraat 15 2513 BH Hague, The Netherlands