

Instructions For Use

RA0664-C-IFU-RUO

Rev. Date: May 7, 2024

Revision: 1

Page 1 of 3

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

Recombinant Double Stranded DNA (dsDNA) (Nuclear Marker); Clone rDSD/8266 (Concentrate)

Availability/Contents: <u>Item #</u> <u>Volume</u>

RA0664-C.1 0.1 ml RA0664-C.5 0.5 ml RA0664-C1 1 ml

Description:

Species: Mouse

Immunogen: Nuclei of Burkitt's cells

Clone: rDSD/8266
Isotype: IgG2b / Kappa
Entrez Gene ID: Not Applicable
Hu Chromosome Loc.: Not Applicable
Synonyms: Not Applicable

Mol. Weight of Antigen: Not Known

Format: 200ug/ml of antibody purified from Bioreactor Concentrate by Protein A/G. Prepared in 10mM

PBS with 0.05% BSA & 0.05% azide.

Specificity: This monoclonal antibody is part of a new panel of reagents, which recognizes subcellular

organelles or compartments of human cells.

Background: These markers may be useful in identification of these organelles in cells, tissues, and

biochemical preparations. This MAb recognizes the double stranded DNA in human cells. It can be used to stain the nuclei in cell or tissue preparations and can be used as a nuclear marker in human cells. This MAb produces a homogeneous staining pattern in the nucleus of normal and malignant cells. Double Stranded deoxyribonucleic acid (ds DNA) is the genetic material of all

cells and many viruses and is a polymer of nucleotides. The monomer consists of

phosphorylated 2-deoxyribose N-glycosidically linked to one of four bases, adenine, cytosine, guanine or thymine. These are linked together by 3',5'-phosphodiester bridges. In the Watson-Crick double-helix model, two complementary strands are wound in a right-handed helix and

held together by hydrogen bonds between complementary base pairs.

Species Reactivity: Human

Positive Control: Jurkat or HeLa cells. Human tonsil or colon, Raji

Cellular Localization: Nucleus

Titer/ Working Dilution: Immunohistochemistry (Frozen and Formalin-fixed): 1-2 μg/ml

Microbiological State: This product is not sterile.

Storage: 2° C 8° C

ScyTek Laboratories, Inc. 205 South 600 West Logan, UT 84321 U.S.A.



Instructions For Use

RA0664-C-IFU-RU

Rev. Date: May 7, 2024

Revision: 1

Page 2 of 3

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

Uses/Limitations: Not to be taken internally.

For Research Use Only.

This product is intended for qualitative immunohistochemistry with normal and neoplastic formalin-fixed, paraffin-embedded


tissue sections, to be viewed by light

microscopy.

Do not use if reagent becomes cloudy. Do not use past expiration date.

Non-Sterile.

Ordering Information and Current Pricing at www.scytek.com



Formalin-fixed, paraffin-embedded human ovarian cancer stained with dsDNA Recombinant Mouse Monoclonal Antibody (rDSD/8266). Inset: PBS instead of primary antibody; secondary only negative control

Procedure:

- Tissue Section Pretreatment (Highly Recommended): Staining of formalin fixed, paraffin embedded tissue 1. sections is significantly enhanced by pretreatment with Tris-EDTA HIER Solution (10x) pH 9.0 (ScyTek catalog# TES500) or Citrate Plus (10x) HIER Solution (ScyTek catalog# CPL500).
- Primary Antibody Incubation Time: We suggest an incubation period of 30 minutes at room temperature. 2. However, depending upon the fixation conditions and the staining system employed, optimal incubation should be determined by the user.
- Visualization: For maximum staining intensity we recommend the "UltraTek HRP Anti-Polyvalent Lab Pack" 3. (ScyTek catalog# UHP125, see IFU for instructions) combined with the "DAB Chromogen/Substrate Bulk Pack (High Contrast)" (ScyTek catalog# ACV500, see IFU for instructions).

Precautions:

Contains Sodium Azide as a preservative (0.09% w/v).

Do not pipette by mouth.

Avoid contact of reagents and specimens with skin and mucous membranes.

Avoid microbial contamination of reagents or increased nonspecific staining may occur.

This product contains no hazardous material at a reportable concentration according to U.S. 29 CFR 1910.1200,

OSHA Hazardous Communication Standard and EC Directive 91/155/EC.

References:

Epstein, A.L. and Clevenger, C.V., Identification of nuclear antigens in human cells by immunofluorescence, immunoelectron microscopy, and immuno-biochemical methods using monoclonal antibodies. In Progress on nonhistone protein research, Vol. 1, Isaac Bekhor, ed., 1985, CRC Press, Boca Raton, FL, pp 117-137

Warranty:

No products or "Instructions For Use (IFU)" are to be construed as a recommendation for use in violation of any patents. We make no representations, warranties or assurances as to the accuracy or completeness of information provided on our IFU or website. Our warranty is limited to the actual price paid for the product. ScyTek Laboratories, Inc. is not liable for any property damage, personal injury, time or effort or economic loss caused by our products. Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining can cause inconsistent results. Variations in fixation and embedding or the inherent nature of the tissue specimen may cause variations in results. Endogenous peroxidase activity or pseudoperoxidase activity in erythrocytes and endogenous biotin may cause non-specific staining depending on detection system used.

Storage: 2° C

ScyTek Laboratories, Inc. 205 South 600 West Logan, UT 84321 U.S.A.



Instructions For Use

RA0664-C-IFU-RUO

Rev. Date: May 7, 2024

Revision: 1

Page 3 of 3

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

