



Recombinant IL3RA / CD123 (Acute Myeloid Leukemia Marker); Clone IL3RA/2947R (Concentrate)

| Availability/Contents: | <u>Item #</u> | <u>Volume</u> |
|-------------------------------|---------------|---------------|
| | RA0673-C.1 | 0.1 ml |
| | RA0673-C.5 | 0.5 ml |
| | RA0673-C1 | 1 ml |

Description:


| | |
|--------------------------|--|
| Species: | Rabbit |
| Immunogen: | Recombinant fragment of human IL3RA protein (around aa 26-171) (exact sequence is proprietary) |
| Clone: | IL3RA/2947R |
| Isotype: | IgG / Kappa |
| Entrez Gene ID: | 3563 |
| Hu Chromosome Loc.: | Xp22.3 or Yp11.3 |
| Synonyms: | Interleukin-3 receptor subunit alpha, CD123; hIL3Ra; IL-3R-alpha; IL-3RA; IL3RAX; IL3RAY; IL3RX; IL3RY; Interleukin-3 Receptor, alpha; Interleukin3 receptor; Interleukin3 receptor, Y-chromosomal |
| Mol. Weight of Antigen: | 70kDa |
| Format: | 200ug/ml of antibody purified from Bioreactor Concentrate by Protein A/G. Prepared in 10mM PBS with 0.05% BSA & 0.05% azide. |
| Specificity: | Recognizes CD123, an interleukin 3 specific subunit of a heterodimeric cytokine receptor. |
| Background: | The receptor is comprised of a ligand specific alpha subunit and a signal transducing beta subunit shared by the receptors for interleukin 3 (IL3), colony stimulating factor 2 (CSF2/GM-CSF), and interleukin 5 (IL5). The binding of this protein to IL3 depends on the beta subunit. The beta subunit is activated by the ligand binding, and is required for the biological activities of IL3. This gene and the gene encoding the colony stimulating factor 2 receptor alpha chain (CSF2RA) form a cytokine receptor gene cluster in a X-Y pseudo-autosomal region on chromosomes X or Y. |
| Species Reactivity: | Human |
| Positive Control: | Lymph Node or Stomach., RPMI-8226 or HDLM-2 cells. Tonsil, THP-1 |
| Cellular Localization: | Membrane |
| Titer/ Working Dilution: | Immunohistochemistry (Frozen and Formalin-fixed): 1-2 µg/ml |
| Microbiological State: | This product is not sterile. |

Storage: 2° C  8° C



ScyTek Laboratories, Inc.
 205 South 600 West
 Logan, UT 84321
 U.S.A.

Uses/Limitations: Not to be taken internally.
For Research Use Only.
This product is intended for qualitative immunohistochemistry with normal and neoplastic formalin-fixed, paraffin-embedded tissue sections, to be viewed by light microscopy.
Do not use if reagent becomes cloudy.
Do not use past expiration date.
Non-Sterile.



Formalin-fixed, paraffin-embedded human Tonsil stained with IL3RA / CD123 Rabbit Recombinant Monoclonal Antibody (IL3RA/2947R).

Ordering Information and Current Pricing at www.scytek.com

Procedure:

- Tissue Section Pretreatment (Highly Recommended):** Staining of formalin fixed, paraffin embedded tissue sections is significantly enhanced by pretreatment with Tris-EDTA HIER Solution (10x) pH 9.0 (ScyTek catalog# TES500) or Citrate Plus (10x) HIER Solution (ScyTek catalog# CPL500).
- Primary Antibody Incubation Time:** We suggest an incubation period of 30 minutes at room temperature. However, depending upon the fixation conditions and the staining system employed, optimal incubation should be determined by the user.
- Visualization:** For maximum staining intensity we recommend the “UltraTek HRP Anti-Polyvalent Lab Pack” (ScyTek catalog# UHP125, see IFU for instructions) combined with the “DAB Chromogen/Substrate Bulk Pack (High Contrast)” (ScyTek catalog# ACV500, see IFU for instructions).


Precautions: Contains Sodium Azide as a preservative (0.09% w/v).
Do not pipette by mouth.
Avoid contact of reagents and specimens with skin and mucous membranes.
Avoid microbial contamination of reagents or increased nonspecific staining may occur.
This product contains no hazardous material at a reportable concentration according to U.S. 29 CFR 1910.1200, OSHA Hazardous Communication Standard and EC Directive 91/155/EC.


References:

- Rollins-Raval M, et al. Appl Immunohistochem Mol Morphol. 2013 May;21(3):212-7

Warranty:

No products or “Instructions For Use (IFU)” are to be construed as a recommendation for use in violation of any patents. We make no representations, warranties or assurances as to the accuracy or completeness of information provided on our IFU or website. Our warranty is limited to the actual price paid for the product. ScyTek Laboratories, Inc. is not liable for any property damage, personal injury, time or effort or economic loss caused by our products. Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining can cause inconsistent results. Variations in fixation and embedding or the inherent nature of the tissue specimen may cause variations in results. Endogenous peroxidase activity or pseudoperoxidase activity in erythrocytes and endogenous biotin may cause non-specific staining depending on detection system used.

Storage: 2° C  8° C



ScyTek Laboratories, Inc.
205 South 600 West
Logan, UT 84321
U.S.A.